

ANTI-TNF DRUGS AND ANTIBODIES (Pilot)

Accreditation Status:	currently not accredited to ISO 17043																										
Date Scheme started:	2024																										
Clinical Applicability:	Assist with drug and antibody monitoring of anti-TNF therapies																										
Analytes:	TNF Adalimumab drug and antibody (SER/065) TNF Infliximab drug and antibody (SER/066)																										
Units for Reporting:	ug/mL (drugs), AU/ml or ng/mL (antibodies)																										
Samples Distributed:	Liquid format. Normal and pathological human serum																										
Number of Distributions per year:	4																										
Number of Samples per Distribution:	1 set of Adalimumab samples (1 x drug, 1 x antibody) 1 set of Infliximab samples (1 x drug, 1 x antibody)																										
Frequency of Distributions:	Every three months as outlined in the Distribution Schedule																										
Schedule of Analysis:	Data entry is via the web for the submission of results. Data analysis is commenced 21 days after sample dispatch. Late returns are accepted and will contribute to the laboratory's cumulative performance statistics																										
Data Analysis:	Reports show method or kit related statistics. Qualitative responses are recorded and assessed in relation to the designated response. All Laboratory Trimmed Mean (ALTM) with truncation at 2SD, SD, and CV%. Reports also show method specific statistics. Individual laboratory performance is expressed in terms of MRBIS, SDBIS, MRVIS and OMIS																										
Performance Scoring:	MRVIS MI scoring based on Consensus Designated Response																										
Criteria of Performance:	Laboratory performance is classified in terms of the MRVIS and OMIS over a Running analytical window of 4 Distributions (12 months) <table><tr><td>Ideal</td><td>MRVIS</td><td><50</td></tr><tr><td>Good</td><td></td><td>50 - 100</td></tr><tr><td>Adequate</td><td></td><td>101 - 200</td></tr><tr><td>Poor</td><td></td><td>>200 or SDBIS >200</td></tr><tr><td></td><td>Total MIS</td><td></td></tr><tr><td>Good</td><td>Zero</td><td></td></tr><tr><td>Adequate</td><td>1-2</td><td></td></tr><tr><td>Poor</td><td>>2</td><td></td></tr></table>			Ideal	MRVIS	<50	Good		50 - 100	Adequate		101 - 200	Poor		>200 or SDBIS >200		Total MIS		Good	Zero		Adequate	1-2		Poor	>2	
Ideal	MRVIS	<50																									
Good		50 - 100																									
Adequate		101 - 200																									
Poor		>200 or SDBIS >200																									
	Total MIS																										
Good	Zero																										
Adequate	1-2																										
Poor	>2																										
Persistent Poor Performance:	Defined as being in the Poor Performance category for two or more successive Distributions																										